



X-ray protective clothing

915 *LytaType* jacket (vest)

Features

- > VELCRO® fastened overlapping front provides all round protection.
- > Front doubles the stated protection at an overlap of approximately 15 cm.

Tailor made

- > Tailor made *LytaType* personal x-ray protective clothing is available. Please contact us for required measurements.

LytaType colours and codes


Please note the colours shown are approximate. A colour swatch is available on request.



Royal
02



Navy
04



Poppy
11



Silver grey
40



Bottle Green
41



S Small for chest sizes (A) 80 – 90 cm girth

| Lead equivalent (Pb)* Front and back | Length (B) | Maximum girth | Product code |
|---|------------|---------------|--------------|
| 0.25 / 0.25 mm | 52 cm | 100 cm | 915/22/S |
| 0.25 / 0.35 mm | 52 cm | " | 915/23/S |
| 0.35 / 0.35 mm | 52 cm | " | 915/33/S |

*Pb: Lead vinyl in accordance with BS EN 61331-1:2014 in the 50 to 150 kV range of x-ray tube voltages.

M Medium for chest sizes (A) 90 – 100 cm girth

| Lead equivalent (Pb)* Front and back | Length (B) | Maximum girth | Product code |
|---|------------|---------------|--------------|
| 0.25 / 0.25 mm | 58 cm | 100 cm | 915/22/M |
| 0.25 / 0.35 mm | 58 cm | " | 915/23/M |
| 0.35 / 0.35 mm | 58 cm | " | 915/33/M |

*Pb: Lead vinyl in accordance with BS EN 61331-1:2014 in the 50 to 150 kV range of x-ray tube voltages.



Protective clothing

Ⓛ Large for chest sizes (A) 100 – 120 cm girth

| Lead equivalent (Pb)* Front and back | Length (B) | Maximum girth | Product code |
|---|------------|---------------|--------------|
| 0.25 / 0.25 mm | 65 cm | 120 cm | 915/22/L |
| 0.25 / 0.35 mm | 65 cm | " | 915/23/L |
| 0.35 / 0.35 mm | 65 cm | " | 915/33/L |

*Pb: Lead vinyl in accordance with BS EN 61331-1:2014 in the 50 to 150 kV range of x-ray tube voltages.

Optional extras: -

| Description | Product code |
|--|--------------|
| Pocket | 900/001 |
| Embroidery in a position to be specified | 9/B20 |

Kenex are certified to Regulation (EU) 2016/425 (PPE) for the design and manufacture of x-ray protective clothing. LytaType aprons and thyroid collars comply with the requirements of BS EN 61331-3:2014 and BS EN 13402-3:2013. They include lead (Pb) vinyl material tested for use with a 50 to 150 kV range of x-ray tube voltages, as specified in BS EN 61331-1:2014.

Notified body: SGS Fimko Oy, P.O. Box 30 (Särkiniementie 3), 00211 Helsinki Finland.
Identification number CE 0598.

