



Protective clothing

# X-ray protective clothing

## 919 *LytaType* kilt (skirt)

### Features

- > Front doubles the stated protection at the overlap.
- > Front overlap of up to 15 cm provides extra central protection. Fitted with VELCRO® and belt fastening.

### Tailor made

- > Tailor made *LytaType* personal x-ray protective clothing is available. Please contact us for required measurements.



Model 919  
Pocket 900/001 shown  
as an optional extra

### *LytaType* colours and codes


Please note the colours shown are approximate.  
A colour swatch is available on request.



Royal  
02



Navy  
04



Poppy  
11



Silver grey  
40



Bottle Green  
41

### **S** Small for waist sizes **(C)** 70 – 84 cm girth

Lead equivalent (Pb)* Front and back	Length <b>(B)</b>	Product code
0.25 / 0.25 mm	55 cm	919/22/S
0.25 / 0.35 mm	55 cm	919/23/S
0.35 / 0.35 mm	55 cm	919/33/S

\*Pb: Lead vinyl in accordance with BS EN 61331-1:2014 in the 50 to 150 kV range of x-ray tube voltages.

### **M** Medium for waist sizes **(C)** 78 – 99 cm girth

Lead equivalent (Pb)* Front and back	Length <b>(B)</b>	Product code
0.25 / 0.25 mm	60 cm	919/22/M
0.25 / 0.35 mm	60 cm	919/23/M
0.35 / 0.35 mm	60 cm	919/33/M

\*Pb: Lead vinyl in accordance with BS EN 61331-1:2014 in the 50 to 150 kV range of x-ray tube voltages.



Protective clothing

**L** Large for waist sizes **(C)** 90 – 115 cm girth

Lead equivalent (Pb)* Front and back	Length <b>(B)</b>	Product code
0.25 / 0.25 mm	65 cm	919/22/L
0.25 / 0.35 mm	65 cm	919/23/L
0.35 / 0.35 mm	65 cm	919/33/L

\*Pb: Lead vinyl in accordance with BS EN 61331-1:2014 in the 50 to 150 kV range of x-ray tube voltages.

**Optional extras: -**

Description	Product code
Pocket	900/001
Embroidery in a position to be specified	9/B20

Kenex are certified to Regulation (EU) 2016/425 (PPE) for the design and manufacture of x-ray protective clothing. LytaType aprons and thyroid collars comply with the requirements of BS EN 61331-3:2014 and BS EN 13402-3:2013. They include lead (Pb) vinyl material tested for use with a 50 to 150 kV range of x-ray tube voltages, as specified in BS EN 61331-1:2014.

Notified body: SGS Fimko Oy, P.O. Box 30 (Särkiniementie 3), 00211 Helsinki Finland.  
Identification number CE 0598.

